



### TEST REPORT TUCHENG

Date: Jun. 15, 2020 Date of Receipt: May. 28, 2020

ORIGINAL

Report No.: TFF9E617 Quantity: 1PC Page Order/Pages: (P1/3) Ref. No.: NIL

Report Title: Aevoe Inc. (Moshi) (T7594) Item: Filter Media

Address: 27F., No. 8, Sec. 5, Zhongxiao E. Rd., Xinyi Dist., Taipei City 110, Taiwan

Test Items		Test Results	Test Methods
Protection Efficiency of Mask(%) (0.075 $\mu$ m NaCl, CMD)	1	99.95	CNS 14755 Z2125-2011 Flow rate: 85.3 (Liter/min)
	2	99.95	
	3	99.95	
	4	99.99	
	5	99.99	
	6	99.99	
	7	99.99	
	8	99.99	
	9	99.99	
	10	99.99	
	Ave.	99.98	

Note: Sample description is given by the client: Nanohedron TM filters

- Note: 1. This report is only responsible for the submitted sample(s), which will be kept for one month period.
2. This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
3. The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

Jui-hung Kao

Director,  
Department of Testing and Certification  
Taiwan Textile Research Institute  
No. 6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108, +886-2-22689839



### TEST REPORT TUCHENG

Date: Jun. 15, 2020 Date of Receipt: May. 28, 2020

ORIGINAL

Report No.: TFF9E617 Quantity: 1PC Page Order/Pages: (P2/3) Ref. No.: NIL

Report Title: Aevoe Inc. (Moshi) (T7594) Item: Filter Media

Address: 27F., No. 8, Sec. 5, Zhongxiao E. Rd., Xinyi Dist., Taipei City 110, Taiwan

Test Items	Test Results	Test Methods
Inhalation Resistance (Pa)	1	212.66
	2	207.76
	3	195.02
	4	196.00
	5	200.90
	6	200.90
	7	192.08
	8	201.88
	9	206.78
	10	217.56
	Ave.	203.15
Exhalation Resistance (Pa)	1	227.36
	2	220.50
	3	215.60
	4	212.66
	5	214.62
	6	215.60
	7	207.76
	8	215.60
	9	221.48
	10	230.30
	Ave	218.15

Note: 1mmH<sub>2</sub>O=9.8Pa.

Note: Sample description is given by the client: Nanohedron TM filters

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

*Jui-hung Kao*

Director,  
Department of Testing and Certification

Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)

Tel : +886-2-22670321 ext. 7107, 7110

Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Jun. 15, 2020 Date of Receipt: May. 28, 2020

ORIGINAL

Report No.: TFF9E617 Quantity: 1PC Page Order/Pages: (P3/3) Ref. No.: NIL

Report Title: Aevoe Inc. (Moshi) (T7594) Item: Filter Media

Address: 27F., No. 8, Sec. 5, Zhongxiao E. Rd., Xinyi Dist., Taipei City 110, Taiwan



Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

*Jui-hung kao*

**Director,  
Department of Testing and  
Certification**

Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)

Tel : +886-2-22670321 ext. 7107, 7110

Fax : +886-2-22675108 , +886-2-22689839